| Kit Specifications | |
|---|---|
| Maximum Protein Input | 200 µg |
| Maximum Column Volume Input | 600 µL |
| Molecular Weight of Recovered Proteins | No molecular weight cut-off |
| Final Endotoxin Levels | ≤ 0.01 EU/µg protein |
| Protein Recovery | 90-95% |
| % Detergent Removal | 90-95% |
| Detergents that can be removed | Including Triton® X-100, CHAPS, NP-40 and Tween 20 |
| Minimum Elution Volume | 50 µL |
| Time to Complete 10 Purifications | 20 minutes |