

Kit Specifications	
Minimum Blood Input	20 µL
Maximum Blood Input	200 µL
Column Binding Capacity	> 50 µg
Average Yield (200 µL of blood)	4-12 µg*
Time to Complete 10 Purifications	30 minutes

*Yield will vary depending on the type of blood processed