Kit Specifications - 96-well	
Minimum Plasma/Serum Input	0.25 mL
Maximum Plasma/Serum Input	2 mL
Size of RNA Purified	All sizes, including microRNA
Time to Complete Purification	< 1 hour