

Kit Specifications	
Minimum Plasma/Serum Input	50 µL
Maximum Plasma/Serum Input	400 µL
Minimum Elution Volume	100 µL
Time to Complete Purifications	< 30 minutes