

Kit Specifications	
Maximum Stool Input	200 mg
Type of Stool Processed	Frozen and fresh stool
Maximum Column Binding Capacity	50 µg
Maximum Column Loading Volume	650 µL
Time to Complete 10 Purifications	30 minutes