

Kit Specifications	
Binding Capacity Per Well	50 µg
Maximum Loading Volume Per Well	500 µL
Size of DNA Purified	All sizes
Maximum Amount of Starting Material: Fungi (Wet weight)	50 mg
Yeast or Gram-positive bacterial culture	0.5 mL - 1 mL
Time to Complete 96 Purifications	40 minutes