

Kit Specifications	
Minimum Urine Input	2 mL
Maximum Urine Input	10 mL
Size of DNA Purified	All sizes of DNA $\geq$ 50 bp
Elution Volume	50-100 μ L
Time to Complete 10 Purifications	40-45 minutes
Average Yields	Variable depending on specimen

*Please check page 6 of the product insert for average yields and the common DNA quantification methods