

Kit Specifications	
Maximum Input Volume	4 mg
Minimum Protein Volume Input	0.25 mg
Maximum Column Volume Input	18 mL
Molecular Weight of Recovered Proteins	No molecular weight cut-off
Final Endotoxin Levels	≤ 0.01 EU/ μ g protein
Protein Recovery	90-95%
% Detergent Removal	90-95%
Detergents that can be removed	Including Triton® X-100, CHAPS, NP-40 and Tween 20
Elution Volume	2-4 mL
Time to Complete 10 Purifications	30-40 minutes