

Kit Specifications	
Minimum Water Input	10 mL
Maximum Water Input	100 mL
Maximum Filter Column Loading Volume	20 mL
Maximum Spin Column Loading Volume	650 μ L
Elution Volume	100 μ L
Time to Complete 10 Purifications	45 minutes