

Kit Specifications	
Minimum Plasma/Serum Input	2 mL
Maximum Plasma/Serum Input	10 mL
Size of DNA Purified	≥ 50 bp
Elution Volume	100 µL
Time to Complete 10 Purifications	15-20 minutes
Average Yields	Variable depending on specimen

Note: Do not exceed the recommended sample input volume of 10 mL.