

<b>Kit Specifications</b>	
Minimum Urine Input	5 mL
Maximum Urine Input	10 mL
Size of RNA Purified	All sizes, including < 200 nt
Time to Complete 10 Purifications	< 30 minutes