

## Technical Specifications

<b>Part No.</b>	<b>Volume</b>	<b>Test Volume</b>	<b>Inaccuracy (+/-)</b>	<b>Imprecision (+/-)</b>
<b>P6-0524</b>	5µl	5µl	1.3%	1.2%
<b>P6-0526</b>	10µl	10µl	0.8%	0.8%
<b>P6-0528</b>	20µl	20µl	0.6%	0.5%
<b>P6-0530</b>	25µl	25µl	0.5%	0.3%
<b>P6-0532</b>	50µl	50µl	0.5%	0.3%
<b>P6-0534</b>	100µl	100µl	0.5%	0.3%
<b>P6-0536</b>	200µl	200µl	0.4%	0.2%
<b>P6-0538</b>	250µl	250µl	0.4%	0.2%
<b>P6-0540</b>	500µl	500µl	0.3%	0.2%
<b>P6-0542</b>	1000µl	1000µl	0.3%	0.2%
<b>P6-0544</b>	2000µl	2000µl	0.3%	0.2%
<b>P6-0548</b>	5000µl	5000µl	0.3%	0.15%